

Message

From: Clemmer, Jill (DPH) [/O=COMMONWEALTH OF MASSACHUSETTS/OU=MASSMAIL-01/CN=RECIPIENTS/CN=JILL.CLEMMER]
Sent: 12/29/2011 1:43:54 PM
To: Servizio, Paul (DPH) [/O=COMMONWEALTH OF MASSACHUSETTS/OU=MassMail-01/cn=Recipients/cn=Paul.Servizio]; Nassif, Julianne (DPH) [/O=COMMONWEALTH OF MASSACHUSETTS/OU=MassMail-01/cn=Recipients/cn=Julianne.Nassif]
CC: Salemi, Charles (DPH) [/O=COMMONWEALTH OF MASSACHUSETTS/OU=MassMail-01/cn=Recipients/cn=Charles.Salemi]; O'Brien, Elisabeth (DPH) [/O=COMMONWEALTH OF MASSACHUSETTS/OU=MassMail-01/cn=Recipients/cn=Elisabeth.O'Brien]; Piro, Peter (DPH) [/O=COMMONWEALTH OF MASSACHUSETTS/OU=MassMail-01/cn=Recipients/cn=Peter.Piro]
Subject: Midazolam, Finally?

Well, they say that life isn't perfect... But we know that Analytical Chemistry can be...

... the exception to the rule is MIDAZOLAM!!!!

Here's a summary of the HPLC-UV, USP results:

Sample ID #	11/30/2011 Midazolam (mg/mL)	12/22/2011 Midazolam (mg/mL)	12/28/2011 Midazolam (mg/mL)	Relative Percent Difference
	0.46	0.453	n/a	2
	1.9	1.97	n/a	4
	3.08	2.82	n/a	9
	0.28	0.275	n/a	2
	0.27	0.255	n/a	6
	2.42	2.44	n/a	1
	5.48	n/a	5.12	7
	5.66	n/a	4.97	13
	5.44	n/a	4.76	13
	5.45	n/a	5.11	6
	5.26	n/a	5.11	3

Recall that the 11/30/11 run had contaminated blanks, because the autosampler needle wash was not on. The re-run of 12/22/11 was incomplete due to a power glitch at 4am. The 12/28/11 re-run "completes" the Midazolam samples (if only), if we accept 111% and 112% QC recovery (instead of 90 – 110%). Can you say UGH!!!!!!!!!!!!!!!!!!!!!!!!!!!!!!!!!!!!?

Jill

PS: I'm done - - burnt-toast.